


FetaLink is a maternal and fetal monitoring system that facilitates the flow of data from medical devices in a manner that supports display needs of clinicians in the acute and ambulatory settings. Alerts for various clinical events may occur and responding to these in a timely manner will aid in patient care.

Responding to Alerts

- Alert parameters define the range of patient data, outside of which an alert will be triggered.
- When the alert fires, the patient chart border will change to **red**, indicating there is an alert.

- The item causing the alert will display with a pink background, and causes an auditory bell sounding the alert.



- The element/component that is alarming is highlighted in pink with a red triangle. There will be a red triangle in the annotation center.
- To acknowledge or annotate an alert, must be signed into FetaLink.



- An indication that not signed in, is the green circle with the checkmark or the acknowledge folder icon is not visible. When logged in, this will illuminate.



- Click the green circle icon to acknowledge the alert without making an annotation.
- Clicking the paper with the paper clip icon, will allow acknowledging the alert and requires an annotation (this is the recommended workflow).



- There are graph bars to indicate the signal strength.
- A red banner and bars indicate a signal strength alert.



- An indicator that the leads may need to be adjusted is the **Heart Coincidence Alert**. The alert is indicated by a waveform with a yellow triangle.








- There may be more than one alert at a time on a patient.
- The border color order for multiple alerts is:
 - Patient alert will be in red.
 - Time Out of Sync is displayed in orange.
 - Device Disconnected is in yellow.
 - Low Signal Quality is displayed in gray.

Patient Specific Alert Parameters

- Select **Alert Settings** from the toolbar in the patient chart.
- This will allow setting alert parameters that are specific to the patient.
- Alerts that can be adjusted are:

- Fetal Heart Parameters
- Maternal Heart Rate Parameters
- Maternal O2 Saturation Parameters
- Maternal Blood Pressure Parameters
- If attempting to set a range that is outside the system allowable range, the cell will turn yellow and will not allow the change to be saved.
- Enter the new desired parameter and select **OK**.
- A **Patients Alert** icon will display beside the room number to indicate the patient now has alert specific parameter(s) set.
 - The icon will only display on the single patient view, it does not display on the central monitor view.
- To remove the patient level alerting, click the **Alert Settings** icon.
- Select **Reset to Default** to remove the patient level alerting.



NOTE: The setting is patient and monitor specific. If patient specific levels are set, they will revert to the default when the patient is disassociated from the monitor for any reason.